

ANNEX II + III: TECHNICAL SPECIFICATIONS + TECHNICAL OFFER

Contract title: IPA 2020 “Supply of equipment for communicable diseases surveillance and for emergency situations”

p 1/58

Publication reference: NEAR/BEG/2022/EA-OP/0219

Columns 1-2 should be completed by the contracting authority

Columns 3-4 should be completed by the tenderer

Column 5 is reserved for the evaluation committee

Annex III - the contractor's technical offer

The tenderers are requested to complete the template on the next pages:

- Column 2 is completed by the contracting authority shows the required specifications (not to be modified by the tenderer),
- Column 3 is to be filled in by the tenderer and must detail what is offered (for example the words ‘compliant’ or ‘yes’ are not sufficient)
- Column 4 allows the tenderer to make comments on its proposed supply and to make eventual references to the documentation

The eventual documentation supplied should clearly indicate (highlight, mark) the models offered and the options included, if any, so that the evaluators can see the exact configuration. Offers that do not permit to identify precisely the models and the specifications may be rejected by the evaluation committee.

The offer must be clear enough to allow the evaluators to make an easy comparison between the requested specifications and the offered specifications.

Unless otherwise specified, the requirements in these Technical Specifications are presented as a minimum standard which the offered goods must meet.

LOT 1 – Reference Microbiology Laboratory

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
1	BECNCHTOP NEXT GENERATION SEQUENCING SYSTEM Quantity: 1			
	<ul style="list-style-type: none"> - Method of sequencing: "sequencing by synthesis" - SBS (sequencing by synthesis). - "Benchtop" instrument (table, not larger than 70x70x70cm and not heavier than 60kg). - Integrated instrument - clonal amplification, sequencing and analysis of data (base calling, alignment, variant calling, reporting) are performed in one instrument. No amplification auxiliary system or additional computer infrastructure is required. -The system should perform paired-read sequencing (bi-directional sequencing) and allow multiple analysis simultaneously of up to 96 samples without the use of PCR reactions. - The throughput of the device (throughput) should be up to 15 GB per one operating cycle. - The instrument should enable the following applications: <ul style="list-style-type: none"> - complete genome resequencing - targeted sequencing - de new sequencing 	<p><i>Manufacturer's name:</i></p> <p><i>Product type, model:</i></p> <p><i>Specifications:</i></p>		

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> - "Mate-pair" sequencing for libraries with inserts of 2-5kb - "Tag-based" gene expression - small RNA sequencing - ribosome profiling - mapping of DNase 1 hypersensitive sites - nucleosome positioning and chromatin structure studies - "ChIP-Seq" examination of sequences involved in protein-RNA interactions - "CNV-Seq" determination of variations in the number of copies by sequencing - "GRO-Seq" determination of the RNA polymerase initiation site - sequencing of ancient DNA samples - "paired-end mRNA" sequencing - Analysis of DNA imprinting and allele specific expression. - Illumination: Light-emitting diodes at 530 nm, 660 nm. -The "Reagent chiller" section should have capacity for the entire reagent cartridge containing cluster-generating reagents, paired-end chemistry, and up to 600 sequencing cycles. Reagents should arrive pre-prepared in an integrated RFID-labelled cartridge. 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<p>-The system should be a single flow cell system. Each "flow cell" is a substrate with one channel that can be read from the top, bottom or both surfaces, depending on the method of preparation. "Flow cells" should be self-positioned, clamped, and coded so that they can be placed in only one correct orientation.</p> <p>-The instrument control computer should be integrated with the sequencer and have a Quad core CPU with 16GB of RAM for instrument control, signal processing, sequence matching, and variant characterization.</p> <p>-The instrument must be delivered with the installed software for primary and secondary data analysis.</p> <p>-Installation, qualification (IQ, OQ and PQ) and training to work on the instrument must be included in the price.</p> <p>-Supply with the instrument a set of reagents for sequencing whole virus genomes and provide training of staff to perform the protocol for at least 5 working days</p>			
2	<p>PRECISE INCUBATION SYSTEM</p> <p>Quantity: 1</p>			
	<p>- High-precision heating system for PCR plates and tubes with heating lid for sample preparation for all NGS platforms;</p> <p>- For NGS sample preparation (NGS prep) and Bead Chip applications;</p>			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> - Sample temperature regulation: $\pm 0.1^{\circ}\text{C}$; - Temperature range: ambient + 5.0°C to 99.0°C; - Ability to select a block of 0.2 ml for PCR plates of 96 wells, 0.6 ml and 1.5 ml tubes, as well as MIDI plates; - Heating lid that minimizes condensation; - Simple temperature calibration; - Dimensions not larger than 21 x 15 x 30 cm; - Weight up to 3 kg 			
3	HIGH SPEED SHAKER Quantity: 1			
	<ul style="list-style-type: none"> -Fast high-performance thermal shaker for handling a wide range of applications adapted for scientific research, with application in the pharmaceutical and biotechnological field. Suitable for shaking micro titer plates, plates with deep wells, tubes and rolls. - Purpose of shaking: micro titer plates, tubes and teats; - Shaking speed - adjustable: 200 - 3,000 rpm; - Constant orbital shaking diameter: 2 mm; - Shake program setting display; - Built-in timer; - Fire-resistant aluminum housing; - Dimensions not larger than cm 80 x 142 x 170 (H x D x W); - Weight up to 2.7 kg; - The instrument should be able to add the following adapters: for micro titer plates, for plates with deep wells and for PCR plates 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
4	HIGH PERFORMANCE MAGNETICS RACK Quantity: 1			
	<ul style="list-style-type: none"> - Magnetic stand designed for product purification during NGS library preparation; - Compatible with 96-well plates and MIDI plates; - Made of ring magnet with integrated Spring Cushion technology, suitable for purifying demanding samples 			
5	GRADIENT THERMAL CYCLER Quantity: 1			
	<ul style="list-style-type: none"> - 7-inch TFT color touch screen, easy to operate; - Fast heating and cooling, precise temperature control, for reliable results; - Instrument can perform conventional PCR, gradient PCR, long-range PCR, isothermal amplification and various experiments - 12 temperature gradients; - Protection against power outages, for your safety and work efficiency; - Flexible options: Tube Mode and Block Mode - Reaction volume: 0-100µl - Module temperature range: 4 ° C-99.9 ° C - Gradient: up to 12 gradients in one reaction - Gradient temperature range: 35 ° C-100 ° C - Temperature range of the heating cover: 40 ° C-110 ° C - Temperature uniformity: ± 0.2 ° C - Temperature accuracy: ± 0.1 ° C - Heating and cooling rates: 5.5 ° C / s 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> - Number of cycles: 99 - Number of steps: unlimited - Temperature increase time: 1s-10min - Storage capacity \geq 1000 programs - Dimensions: 260mm x 400mm x 260mm, 11kg - Noise at work: <55dB 			
6	REAL TIME PCR INSTRUMENT Quantity: 1			
	<ul style="list-style-type: none"> -96 wells, 6 channels -Reaction volume: 0-100μL -Temperature range: 0-100 ° C -Heating / cooling method: Peltier -Maximum heating rate: 6.1 ° C / sec -Maximum cooling rate: 5.0 ° C / sec -Temperature accuracy: \pm 0.1 ° C -Temperature uniformity: \pm 0.1 ° C -Gradient block: 12 row, gradient range: 1-40 ° C, -Detection time: 7 sec for 96 wells for all channels -Possibility of detecting 1 copy -The real-time PCR device should have the following excitation range: <ul style="list-style-type: none"> - 465/510 (FAM, SYBR Green, EVA Green, LC Green); 527/568 (HEX, VIC, TET, JOE); 580/616 (ROX, Texas Red); 632/664 (Cy5); 680/730 (Alexa Fluor680); 465/616 (FRET) -Probe: TaqMan, Molecular beacons, Scorpion, FRET - Real-time PCR should be an open system for consumables - Real-time PCR device suitable for the following analyzes: quantitative analysis, absolute 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	quantification, relative quantification, genotyping analysis, SNP analysis, dissociation curve analysis, endpoint analysis, HRM - Dimensions of Real-time PCR devices: 355mm x 480mm x 485mm, 30kg -Real-time PCR operating environment: temperature: 10 ° C-30 ° C, humidity: 20% -80%			
7	MICROCENTRIFUGE Quantity: 2			
	-Maximum rotor speed 14800 rpm -Maximum g-force: 16163 xg -Maximum capacity 12 x 1.5/2 ml tube -Timer setting: 20 sec to 99 min 59 sec in 1 sec increments -Acceleration / deceleration time from max speed: 17 sec / 18 sec -Pulse mode for short, fast centrifugation -Dimensions not larger than (H x L x W): 180mm x 280mm x 230mm -Device weight maximum 6.4 kg -Noise level at 1m distance <60 dB -Rotor imbalance detection -Rotor exchange without tool -Rotor with protective cover			
8	UNIVERSAL CENTRIFUGE WITH ROTORS Quantity: 2			
	-Rotor speed 16000 rpm -Maximum g-force: 23511 xg -Maximum capacity 4 x 400 ml tubes			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Speed adjustment: up to 16000 rpm / 23511 xg, in divisions of 100 rpm or 100 xg -Timer setting: 9 min 59 sec in 1 sec increments -Unlimited continuous centrifugation mode -Pulse mode for short centrifuges at maximum speed -10 levels of acceleration and 10 levels of deceleration -Pulse mode for short, fast centrifugation -Audible signal to complete spinning and report a fault, with the possibility of switching off if necessary -Possibility to change set parameters (speed, RCF, time, acceleration / deceleration) during centrifuge operation -Dimensions not larger than (H x L x W): 355mm x 460mm x 550mm -Device weight maximum 48 kg -Noise level at 1m distance <68 dB -Automatic door lock after start -Aluminum rotor with airtight cover, capacity 24 seats for 2 ml tubes, fixed rotor angle 45 °, k factor 185 -Aluminum rotor with a capacity of 10 places for 10 ml tubes, fixed rotor angle 35 °, k factor 500 - Aluminum fixed angle rotor for 15 ml conical and round-bottom tubes k factor 500 -Swingout rotor 0-90 ° for 2 x micro titer plates, maximum RPM 4700, maximum force 2721 RPM" 			
9	UNIVERSAL CENTRIFUGE WITH ROTOR FOR TUBES			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	Quantity: 1			
	<ul style="list-style-type: none"> -Benchtop , small centrifuge -Max. speed: 6800 - 8000 rpm -Max. RCF: 6153 xg -Max. capacity: 8 x 15 ml tube -Voltage 220V -Angle rotor 8 -place -Benchtop model, small size -Microprocessor with large LCD display -Electrical lid lock -Air Cooling System -Control panel with Touch-Operation 			
10	MINISPIN CENTRIFUGE			
	Quantity: 4			
	<ul style="list-style-type: none"> -Capacity 8x 2.0ml -Centrifugal force 2000 xg -Fix speed 6000 rpm -Power-ON time 99 min -Timer display 7 segment LED -Quick stop -Cover opening automatic -Rotor exchange without tool -Protective cover -Noise level 48 dB -Dimensions (W x H x D) 155 x 105 x 175 mm -Weight 1.4 kg -Protection class according to DIN EN 60529 IP 30 			
11	ANTIVIBRATION TABLE			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	Quantity: 1			
	-Anti-vibration table without wheels -The bench must withstand the weight of the instrument - more than 100 kg -Table dimensions: up to 122cm x 92cm x 76cm			
12	LIQUID HANDLING SYSTEM FOR PREPARATION OF NGS LIBRARIES Quantity: 1			
	-Min 25 positions on the deck - the possibility of logical positioning according to the requirements of the protocol - Infrared curtain for security -Rotating gripper with unique offset finger design - Internal LED light - Bright, multiple color- and pattern-coded status light bar - Spacious, open-platform with access from all sides - Onboard cameras for live broadcast and on-error video capturing - Linear motion control - Optional Fly-by barcode reader - Orbital shaker, tip washer and magnetic plate included - The pipetting liquid displacement mechanism prevents any type of cross-contamination - Pipette tips can be automatically changed between each aspiration if required - The software will generate a warning sign if the aspirated volume is greater than the volume of the pipette tip filled on the mandrel			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> - Pipettes can move independently in X,Y, Z and D axes - percentage error depending on volume $\pm 1-5\%$ - 8 independent pipetting channels - Pipetting volume 0.5 - 5000 μl - Dimensions: Width: not more than 112 cm Depth: not more than 81 cm Height: not more than 104 cm - Weight: not more than 146 kg 			
13	WATER BATH Quantity: 1			
	<ul style="list-style-type: none"> - TFT display - Useful volume 20.3 L - Thermal insulation cover made of stainless steel, with double walls with internal curvature 			
14	ANALITICAL BALANCE Quantity: 1			
	<ul style="list-style-type: none"> - Range: 0-220 g - Readability: 0.1 mg - Repeatability: 0.2 mg - Linearity ± 0.3 mg - Weighing surface 120 mm 			
15	PCR BOX WITH UV LAMP AND HEPA FILTER Quantity: 1			
	<ul style="list-style-type: none"> -External dimensions: 650 mm (w) x 560 mm (d) x 875 mm (h) -Internal dimensions: 640 mm (w) x 540 mm (d) 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Weight: 60 kg -Alarm system: Low airflow - Warning light and audible alarm -Operator safety: Safety screen with constant monitoring of UV and fluorescent light status -Laminar air flow: vertical flow 0.4 m / sec -Filtration: HEPA filter for air of purity ISO Class 5 -UV light: 20 V UV germicidal lamp -UV control: 4 preset time intervals with LED status indication -UV tube lifetime alarm: Light warnings and audible alarm after 1000 hours -Interior lighting: Full width LED lighting -Electrical requirements: 230V / 50Hz -Noise level: <55db (A) -Construction: Epoxy-coated steel cabinet to prevent corrosion. Polycarbonate side and front protection panels. -Stainless steel work surface 			
16	FRAGMENT ANALYZER			
	<p>Quantity: 1</p> <ul style="list-style-type: none"> -Detection method: Fluorescence -Light source: LED -Connection: USB -Minimum power: 30W -Weight: 15 kg -Dimensions: 38x30x40 cm -Automated sampling: 1-96 samples 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Consumable gel cartridge: 100-300 samples, single channel cartridge, DNA: S1, S2, S3, F3, N1, N3, RNA: R1, NR1, Protein: P2 -Quick analysis: 1-7 minutes per sample (below 1 kb) -Resolution: 1-4 bp (below 500 bp) -Sensitivity: 5 pg / µl, 1 pg / µl (if diluted with distilled water) -Minimum sample volume: 1 µl (Micro vials. C104250), 10 µl (0.1 ml tube), 20 µl (0.2 ml tube) -Sample consumption: ≤1pL -Software - Relative qualitative and quantitative analysis, Electrophoresis, Analysis of images, Forming reports in accordance with the needs of users 			
17	BIOLOGICAL SAFETY CABINET CLASS II Quantity: 2			
	<ul style="list-style-type: none"> -Class II microbiological safety cabinet -Device in accordance with European standard for microbiological safety cabinets EN 12469: 2000 -Electronic closing and opening of the front panel using an electric motor -Infrared sensor for control and protection when closing the front movable panel -Control unit with a touch screen of at least 4.3"image quality 130,000 pixels or more -Measurement and display of the current flow rate on the screen in m / s -Automatic airflow compensation to maintain set point -Audible and visual alarm 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Display of fan operating time, cabinet operation and service interval on the screen -PIN code-protected device specifications -Possibility of on-screen display: date of last and next service, date of filter installation and filter label -Air purity: ISO Class 4 (Class 10) in accordance with EN ISO 14644-1 (1999) -Main filter: H14 HEPA filter - 99.97% efficiency for particles > 0.3µm -Primary discharge filter: H14 HEPA filter - 99.97% efficiency for particles > 0.3µm -Additional outlet filter: H14 HEPA filter - 99.97% efficiency for particles > 0.3µm -Work surface made of stainless steel SS 316 -External dimensions of the cabinet without stand, ± 5% (W x D x H): 1200mm x 750mm x 1330mm -Workspace dimensions, ± 5% (W x D x H): 1107mm x 510mm x 741mm -Supplied with additional stand -At least two AC power outlets in the cabinet workspace that are activated and deactivated by pressing the display key -LED interior lighting with adjustable light intensity -Cabinet to be equipped with UV germicidal lamp -Possibility to adjust the duration of sterilization with a UV lamp and delayed start of sterilization 			
18	AUTOMATED SYSTEM FOR EXTRACTION OF NUCLEIC ACIDS Quantity: 2			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> - Extraction according to the principle of magnetic beads - 1-96 samples per cycle - 6 modules for plates, each plate for 16 samples - Plates: 2 x 8 samples - The possibility of using pre-filled whales for extraction - Duration of extraction for 96 samples: 24 minutes - Reaction volume: 30-1000µL - Sample volume: 200µL - Rotation speed: ≤2000rpm - Range of temperature control: Pyrolysis heating: room temperature-120 ° C, Elution heating: room temperature-120 ° C - Areas of application: clinical diagnostics, forensics, research, food quality control, control of the epidemiological status of the area . This instrument can process serum, plasma, whole blood, swabs, amniotic fluid, feces, tissues, paraffin sections, bacteria, fungi and other types of samples - Built-in 7-inch LCD display, touch screen - Flexible creation, deletion and modification of extraction programs - Possibility of storing 10,000 protocols, USB port - Extraction efficiency: high reaction efficiency for nucleic acids ≥95% - Contamination control: HEPA exhaust filter module with negative pressure and built-in UV lamp - Pre-installed program for UV disinfection - Difference between wells CV≤1% 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> - Temperature range: between 10 ° C and 30 ° C - Humidity ≤85% - Operating noise <60 dB - Dimensions: 490 * 520 * 480mm 			
19	FLUOROMETER Quantity: 1			
	<ul style="list-style-type: none"> - More sensitive than quantification based on UV absorption, making it ideal for valuable samples - Accurately quantifies DNA, RNA and protein in < 3 seconds per sample - The new integrated reagent calculator reports the amount of color and buffer required - Use only 1 mL of sample - Flexible options for exporting results: Wi-Fi, USB drive or direct connection with USB cable - Dimensions: 13.6 cm x 25 cm x 5.5 cm, 743g - Excitation filters: Blue 430–495 nm, Red 600–645 nm - Emission filters: Green 510–580 nm, Red 665–720 nm - Processing time: ≤5s / sample - Warm-up time: <35s 			
20	NGS ANALYSIS SOFTWARE (PC included) Quantity: 1			
	<ul style="list-style-type: none"> -Microbial Genomics For Outbreak-, Resistom-, Virulome-Investigation, and Real-time Surveillance -Allows automatic processing and analyzing of NGS (e.g., Illumina, Ion Torrent, or 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<p>PacBio) and Sanger capillary electrophoresis sequence data</p> <ul style="list-style-type: none"> -User friendly-No scripting or bioinformatics skills needed to process your data and run your analyses. -Possibility to download pre-defined typing schemes or quickly create a new scheme based on a reference genome or allele libraries. -Possibility to setup an assembling and processing pipeline to analyze hundreds of samples without any further user intervention -Assemble raw reads (FASTQ) with integrated SKESA, Velvet, SPAdes, and BWA algorithms -Read and analyze assembly files (ACE, BAM, FASTA) -Contamination check with Mash Screen -Integrated database with an ability to store, search, retrieve, export, and create reports from experiment, epidemiologic, and DNA sequence data -Data fields are compliant with the meta-data requirements of the EBI European Nucleotide Archive (ENA) -Manage and backup all data (sequence and epi-data) -Possibility to download from NCBI complete/draft genomes or SRA reads 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Possibility to compare new sequence entries against stored data -Automatically cluster alerts of possible outbreaks -Possibility for rapid and easy share typing schemes with others -Possibility to request a non-public epidemiologic data server for a closed group of users -Encryption (SSL) of all data in transmission - 1 PC: Computer Specification Microsoft Windows 64-bit or Linux 64-bit, quad-core processor, 16-32 GB RAM, 50 GB hard disk (HD), and Internet connection (via ports 80 & 443) 			
21	LABORATORY WATER DEMINERALIZATION SYSTEM Quantity:1			
	<ul style="list-style-type: none"> -Resistivity: 18.2 MΩ.cm at 25°C >1MΩ.cm @ 25°C -Flow rate: Up to 1.2 l/min -Type 1, Type 2, Type 3 -Recommended volume: Up to 10l/day1 Type 1, Up to 10l/day1 Type 2, Up to 30l/day Type3 -TOC value: < 5 ppb2 Type 1 <50 ppb Type 2 <200 ppb3, Type 3 -Bacteria TVC: <0.1cfu/ml4 Type 1 <100 cfu/ml Type 2 <1000 cfu/ml3 Type 3 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Endotoxin: <0.001EU/ml⁵ -RNases: <1pg/ml⁶ -DNases: <5pg/ml⁶ -Particulates: 0.2µm filtration -Application includes molecular biology e.g. DNA sequencing and PCR 			
22	ULTRA LOW FREEZER (minus 40 to minus 90°C) upright, volume 700-800 L Quantity:2			
	<ul style="list-style-type: none"> -Device temperature range from -40 ° C to -90 ° C -Net device volume from 700L to 800L -Max external dimensions of the device (Width x Depth x Height): 1030 x 1040 x 1993 mm -Microprocessor control of the device Pt-1000 temperature sensor -Color touchscreen LCD for parameter control -Hydrocarbon-based coolant for environmental protection -Insulation of polyurethane foam and vacuum panels with a maximum thickness of 82 mm -Exterior doors with insulation -Security lock with the possibility of locking with a key or NFC card -Stainless steel or aluminum Interior door with handle, at least 2 pieces 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Automatic valve for vacuum release opening the door in case of emergency -Visual and audio alarm system for power failure, high temperature, low temperature, open door -Ability to protect device parameter settings with password -Possibility of graphical display of stored temperatures -Noise level <52 dBA -Electricity consumption 11.9 kWh / 24h -The device is equipped with software for recording, monitoring and visualization of temperature data and lists of events and alarms -Graphical display of temperature, upper and lower limits of temperature alarms depending on the time, on the screen of the device -Ability to transfer all recorded temperature data and lists of events and alarms up to 3 months back, to USB memory or SD card -With trays 2 inch 24 x SDR-624-P 6x4 576 139.45 565.4 339.59 134 137 54,5 -24 pcs stainless steel racks with trays for 2" boxes, total capacity of at least 576 boxes 2" 			
23	ULTRA LOW FREEZER (minus 86°C) upright, volume 50 L Quantity:1			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	-External dimensions 825x595x645 mm -Dimension inner 630x435x428 mm -Weight 87 kg -Material inner cabinet Stainless steel -Material outer cabinet Stainless steel -Insulation type Polyurethane with cyclopentane -Insulation thickness 80 mm -Mobility Feet -Volume min 90 L -Fitures: Door Lock, LED light, Battery backup			
24	MALDI (Matrix-Assisted Laser Desorption/Ionization) APPARATUS Quantity: 1			
	-Benchtop / table mass spectrometer of linear configuration with MALDI-TOF technology for microbiological identification -Operation in the mass range up to at least 500,000 Da -Minimum laser life: 400 - 500 million shots -200Hz laser enables fast sample processing; min 300 - 400 samples / h -Data system, monitor, barcode reader -Ability to connect to the laboratory information system independently of the manufacturer			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Ability to connect to an antibiotic susceptibility testing system independent of the system manufacturer -Possibility of connecting to the system remotely -Possible export of data in a form that allows data processing with other programs -Includes IVD reference library with profiles of microorganisms that allows identification of the following groups of microorganisms: enterobacteria, gram negative nonfermentative bacteria, gram positive cocci, gram positive bacilli, anaerobes, legionella, nocardia, campylobacteria, HACEK bacteria, yeasts and broad species environment; a total of a minimum of 3800 species. -Includes a reference library with profiles of microorganisms that allows the identification of molds, and a special protocol for the cultivation and preparation of samples, a minimum of 240 species -Includes an IVD reference library that allows the identification of third level biosecurity microorganisms. -IVD Module for optimized analysis of direct stars from positive blood cultures -IVD Module for the module for automatic subtyping of certain species (KPC-producing <i>Klebsiella pneumoniae</i> and <i>Escherichia coli</i>; <i>Bacteroides fragilis</i> <i>cfiA</i> subtyping) 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -IVD Module for detecting the presence of beta-lactamases, carbapenemases -The matrix can be applied within 30 minutes from the preparation of the sample smear on the tiles -Possibility of interruption of measurements, and change of sample plates in case of emergency samples -Reusable tiles (washable) -Possibility of using disposable plates with the possibility of using all positions regardless of the number of inserts of the plate in the instrument -Start-up reagent set included at least 2000 samples and 200 positive blood culture samples -Built-in diaphragm pump system allows easy maintenance with minimum service interval (3 years) -Ionization source self-cleaning protocol without vacuum exclusion -Appropriate CE-IVD certificate for equipment 			
25	MICROSCOPE w LED FLUORESCENCE AND CAMERA Quantity:1			
	-Robust metal stand			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Binocular ergophoto tube 100%:0% and 0%:100% with inclination max 20°, adjustable height of viewing for min. 44mm -Objective nosepiece, 6 position, encoded, with functions: <ul style="list-style-type: none"> - light manager (automated adjusting - - light intensity in conjunction with objectives) - automated measuring ratio - Mechanical table with drive which can be adjust for 15mm -Two integrated ergonomically positioned snap image buttons on left and right side of microscope (The snap image buttons allows to acquire images and videos directly on USB, without PC) -Achromatic aplanatic condenser 0,9 for magnification 1- 100x with 5 positions for phase contrast and dry darkfield -Illumination LED min 10W encoded, full Koeller, adjusting intensity on stand and ECO mode (automatically goes to stand by after being idle for 15 min.) -Active light manager with an adjustable light intensity, suitable for all kinds of objectives -Light intensity is memorized per objective and provides uniform brightness at all magnifications, eliminating manual lamp intensity adjustments when changing objectives 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Objective: 10x/0.3, 20x/0.5, 40x/0.75, 63x/0.85, and 100x/1.3Oil -Eyepiece10x with field of view 23 mm -Additional magnification 1,25x and 1.6x -LED fluorescence with 4 wavelengths (385nm, 470nm, 565nm and 625nm) with power supply and control directly from microscope -Filter set for florochrome (DAPI, Alexa Fluor 405, FITC, eGFP, TRITC, Cy3, AlexaFluor 633, Cy5) -Digital microscope monochromatic camera with HD resolution, live image min 30fps via HDMI, USB 3.0) -Image sensor size min. 13mm, pixel size min 5.86µm, exposition time 2s or more, same producer as microscope -Appropriate adapter and software for multichannel fluorescence, extended focus, measurement) 			
26	MICROSCOPE w BRIGHTFIELD,PHASE CONTRAST, DIGITAL CAMERA AND SOFTWARE Quantity:1			
	<ul style="list-style-type: none"> -Robust metal stand -Binocular photo tube 50%:50% with inclination 30°, adjustable height of viewing for min 35mm 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<p>-Objective nosepiece, 5 position, encoded, with functions:</p> <ul style="list-style-type: none"> • light manager (automated adjusting light intensity in conjunction with objectives) • automated measuring ratio <p>-Mechanical table with drive which can be adjust for 15mm</p> <p>-Two integrated ergonomically positioned snap image buttons on left and right side of microscope (The snap image buttons allows to acquire images and videos directly on USB, without PC)</p> <p>-Achromatic aplanatic condenser 0,9 for magnification 1- 100x with 5 positions for phase contrast and dry darkfield</p> <p>-Illumination LED min 10W encoded and halogen min 12V 35W, full Koeller, adjusting intensity on stand and ECO mode (automatically goes to stand by after being idle for 15 min.)</p> <p>-Active light manager with an adjustable light intensity, suitable for all kinds of objectives</p> <p>-Light intensity is memorized per objective and provides uniform brightness at all magnifications, eliminating manual lamp intensity adjustments when changing objectives</p> <p>-Objective: 10x/0.25Ph1, 20x/0.45Ph2, 40x/0.65Ph2, 63x/0.85Ph3 and 100x/1.25Ph3Oil</p>			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	-Eyepiece 10x with field of view 23 mm -Digital microscope colour camera with Ultra HD/4K resolution, live image min 30fps via HDMI, USB 3.0) -Image sensor size min. 8mm, pixel size min 1.8µm, exposition time min. 1s, same producer as microscope -Appropriate adapter and software			
27	MICROSCOPE w ULTRA DARK FIELD AND DIGITAL MICROSCOPE COLOR CAMERA Quantity:1			
	-Robust metal stand -Binocular photo tube 50%:50% with inclination 30°, adjustable height of viewing for min 35mm -Objective nosepiece, 6 position, encoded, with minimum following functions: <ul style="list-style-type: none"> • light manager (automated adjusting light intensity in conjunction with objectives) • automated measuring ratio -Mechanical stage with travel range min. 105x85 mm for microtiter plates -Two integrated ergonomically positioned snap image buttons on left and right side of microscope (The snap image buttons allows to acquire images and videos directly on USB, without PC) -Ultra darkfield condensor 1.2/1.4 (0.75-1.0)			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Illumination LED min 10W encoded , full Koeller, adjusting intensity on stand and ECO mode (automatically goes to stand by after being idle for 15 min.) -Active light manager with an adjustable light intensity, suitable for all kinds of objectives -Light intensity is memorized per objective and provides uniform brightness at all magnifications, by eliminating manual lamp intensity adjustments when changing objectives -Objective: 10x/0.2 with FWD min.11mm, 20x/0.40 with FWD min 7.2mm -Eyepiece10x with field of view 23 mm -Digital microscope colour camera with Ultra HD/4K resolution, live image min 30fps via HDMI, USB 3.0) -Image sensor size min. 8mm, pixel size min 1.8µm, exposition time min. 1s -Same producer as microscope -Appropriate adapter and software 			
28	MICROSCOPE w BRIGHT FIELD Quantity:4			
	<ul style="list-style-type: none"> -Robust metal stand -Binocular photo tube 50%:50% with inclination 30°, adjustable height of viewing for min 35mm -Objective nosepiece, 5 position, encoded, with functions: 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> light manager (automated adjusting light intensity in conjunction with objectives) automated measuring ratio -Mechanical table with drive which can be adjust for 15mm -Two integrated ergonomically positioned snap image buttons on left and right side of microscope (The snap image buttons allows to acquire images and videos directly on USB, without PC) -Achromatic aplanatic condenser 0,9 for magnification 1- 100x -Illumination LED min 10W encoded and halogen min 12V 35W, full Koeller, adjusting intensity on stand and ECO mode (automatically goes to stand by after being idle for 15 min.) -Active light manager with an adjustable light intensity, suitable for all kinds of objectives -Light intensity is memorized per objective and provides uniform brightness at all magnifications, eliminating manual lamp intensity adjustments when changing objectives -Objective: 10x/0.25, 20x/0.45, 40x/0.65, 63x/0.85 and 100x/1.25 Oil -Eyepiece 10x with field of view 23 mm			
29	MICROSCOPE with BRIGHT FIELD and DIGITAL CAMERA			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	Quantity: 5			
	<p>-Robust metal stand</p> <p>-Binocular photo tube 50%:50% with inclination 30°, adjustable height of viewing for min 35mm</p> <p>-Objective nosepiece, 6 position, encoded, with functions:</p> <ul style="list-style-type: none"> • light manager (automated adjusting light intensity in conjunction with objectives) • automated measuring ratio <p>-Mechanical table with drive which can be adjust for 15mm</p> <p>-Two integrated ergonomically positioned snap image buttons on left and right side of microscope (The snap image buttons allows to acquire images and videos directly on USB, without PC)</p> <p>-Achromatic aplanatic condenser 0,9 for magnification 1- 100x</p> <p>-Illumination LED min 10W encoded, full Koeller, adjusting intensity on stand and ECO mode (automatically goes to stand by after being idle for 15 min.)</p> <p>-Active light manager with an adjustable light intensity, suitable for all kinds of objectives</p> <p>-Light intensity is memorized per objective and provides uniform brightness at all magnifications, eliminating manual lamp intensity adjustments when changing objectives</p>			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Objective: 10x/0.25, 20x/0.45, 40x/0.65, 63x/0.85, 100x/0.8 and 100x/1.25Oil -Eyepiece10x with field of view 23 mm -Additional magnification 1,25x and 1.6x -Digital microscope colour camera with Ultra HD/4K resolution, live image min 30fps via HDMI, USB 3.0) -Image sensor size min. 8mm, pixel size min 1.8µm, exposition time min. 1s, same producer as microscope. -Appropriate adapter and software 			
30	STEREO MICROSCOPE Quantity: 1			
	<ul style="list-style-type: none"> -Apochromatic optics -Min. 11° or more Greenough optical system -Zoom ratio 8:1 -Magnification 6,3x...50x with objective 1x, -Free working distance min 92mm -Compact stand with LED transmitted illumination: brightfield, darkfield and oblique illumination -Reflected LED segmented ring illumination with power supply from stand -Integrated binocular photo tube with inclination not more than 35° -Eyepiece 10x with field of view min 23 			
31	AUTOMATED ELISA READER Quantity: 1			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Fully automated ELISA processing for low to medium sample volume with minimal manual input -Open system -High reliability and traceability due to automatic identification of barcodes of patient samples and ready-to-use reagents -Fast processing of up to 50 tests per hour -Capacity for up to 3 plates and 144 samples per run -Convenient operation of the software including scanning of QC certificates using a 2D-hand barcode scanner -Sample types Plasma, serum, CSF and stool -Parameter capacity per run up to 36 -Shakeable incubators 2 -Sample tubes Outer diameter 10 – 16 mm, height up to 100 mm -User interface Microsoft ® Windows ® 7 or similar -Online connection ASTM interface, bi-directional -Dimensions (WxDxH): up to 1600 mm x 750 mm x 1150 mm (incl. attached All-in-one PC) -Weight: up to 100 kg 			
32	ELISA READER			
	Quantity: 2			
	-Detection modes - UV/vis absorbance spectra (Wavelength range 220-1000 nm)			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<p>-Measurement modes -Endpoint and kinetic; Spectral scanning (absorbance); Well scanning</p> <p>-Microplate formats - Possibility to upgrade the option for reading plates from 6-1536 wells</p> <p>-Light sources - High energy xenon flash lamp</p> <p>-Detectors - CCD spectrometer</p> <p>-Read times - less than 1 sec/well</p> <p>-Sensitivity - Full spectrum captured in < 1 s/well</p> <p>-Selectable spectral resolution: 1, 2, 5, and 10 nm</p> <p>-OD range: 0 to 4 OD</p> <p>-Accuracy: < 1% at 2 OD</p> <p>-Precision: < 0.5% at 1 OD and < 0.8% at 2 OD</p> <p>-Shaking - Linear, orbital, and double-orbital with user-defi nable time and speed</p> <p>-Incubation - +3°C above ambient up to 45 °C (the upper heating plate is always 0.5 °C warmer than the lower one in order to avoid condensation)</p> <p>In accordance with FDA 21 CFR Part 11</p> <p>-Software included in the price</p> <p>-Computer included in the price</p> <p>-Optional - small volume plate with 16 wells (2 µL)</p>			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
33	ELISA WASHER Quantity: 2			
	<ul style="list-style-type: none"> -3.5-inch LCD display. -100 wash programs. -Aspirating in 2 positions, residual volume <0.7 µl. -Multi-plate can be selected by default settings or manual operation. -Special rinse procedure makes easy to maintain the device and reduces contamination between plates. -Automatical rinse procedure by distilled water when turn on or turn off can avoid crystallizing in the pipes. -Suitable for plate models: flat-, round -, U-, V- plate (96-well and 48-well) -Manifold types: 8- and 12- way -Wash times: 1 to 99 times adjustable -Wash lines: 1-12 lines selectable -Wash mode: plates or strips -Dispensing volume: 50-950 µl, interval 50 µl adjustable (May extend to 0-6000 µl, interval 25 µl adjustable) -Number of dispensing channels: 3 channels -Soaking and shaking time: 0-999 (seconds/minutes/hours) adjustable -Aspiration time: 0.1-9.9 seconds adjustable, interval 0.1second -Pipeline wash time: 0-240 seconds adjustable 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	-Distilled water automatically wash interval: 0-20 plates - To be supplied with 8 (eight) bottles for washer/waste			
34	ULTRASONIC BATH Quantity: 2			
	-Ultrasonic frequency 40 KHz -Tank material Stainless steel Cr-Ni 18/10 -Timer 1 - 30 minutes with automatic shutdown -Tank capacity 12 L -Valve for draining liquid from the tank, with hose connection -Total power 1320W -Power of ultrasonic generator 720W -Heater power 600W -Heating control Adjustable up to 90°C with automatic shut-off -Optimal liquid temperature 40 °C -Internal dimensions 300mm × 240mm × 200mm (L × W × H) -External dimensions 310mm × 161mm × 350mm (L × W × H)			
35	INCUBATION SYSTEM (minimal volume 100 L) Quantity: 8			
	-Minimum 100 L			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Upper temperature range: Up to 80 °C or better -Load per rack/shelf: minimum 15 kg -Audio/Visual alarm -Inner glass door -Temperature fluctuation: maximum $\pm 0,2$ °C (at 37 °C) Temperature range: +5 °C above ambient temperature to +100 °C -High temperature accuracy thanks to APT.line™ technology or similar -Natural convection -Controller with LCD display -Electromechanical control of the exhaust air flap -Inner door made of tempered safety glass 2 chrome-plated racks -Stackable -Class 3.1 integrated independent temperature safety device (DIN 12880) with visual alarm -Ergonomic handle design -USB port for recording data 			
36	INCUBATION SYSTEM (minimal volume 200 L) Quantity: 5			
	<ul style="list-style-type: none"> -Minimum volume 200 L -Temperature range: +5 °C above ambient temperature to +100 °C -High temperature accuracy, APT line technology or similar 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Forced convection -Controller with LCD display -Electromechanical control of the exhaust air flap -Inner door made of tempered safety glass -2 chrome-plated racks -Class 3.1 integrated independent temperature safety device (DIN 12880) with visual alarm -Ergonomic handle design -USB port for recording data 			
37	INCUBATION SYSTEM WITH CO2 OPTION (minimal volume 50 L) Quantity: 3			
	<ul style="list-style-type: none"> -Minimum volume 50 L -Temperature Control Range & Fluctuation: °C at +5 ~ +50, ±0.1 -Temperature Uniformity: ±0.25 °C -CO2 Control Range & Fluctuation: % 0 ~ 20, ±0.15 -Humidity Level & Fluctuation: %RH 95, ±5 -Sterilisation Method: H2O2 -Decontamination (optional) -Temperature Sensor: Thermistor -CO2 Sensor: Dual IR -Display: LCD Touch Screen -Exterior Material: Painted Steel (rear cover not painted) -Interior Material: Stainless Steel Copper Enriched Alloy 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	-Insulation Material: Extruded polystyrene -Heating Method: Direct Heat & Air Jacket System			
38	INCUBATION SYSTEM WITH COOLING OPTION Quantity: 4			
	-Temperature range: 10 -60 C or better -Interior volume minimum: 53L -Load per rack/shelf: minimum 15 kg -Optional stacking -Inner glass door -Audio/visual alarm -Temperature variation/uniformity: maximum $\pm 0,5$ C (at 37 C) - Display: LCD Touch Screen			
39	MULTIGAS INCUBATION SYSTEM Quantity: 1			
	- External Dimensions (W x D x H): up to 480 x 550 x 585 mm ($\pm 5\%$) - Volume 50 liters or more - Temperature Control Range & Fluctuation: at $+5^{\circ}\text{C} \sim +50$, ± 0.1 - Temperature Uniformity: $\pm 0.25^{\circ}\text{C}$ - CO ₂ Control Range & Fluctuation: 0% ~ 20%, $\pm 0.15\%$ - O ₂ Control Range & Fluctuation: 1% -18% and 22% - 80%, ± 0.2			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> - Humidity Level & Fluctuation: 95% RH, ± 5 - CO2 Sensor: Dual IR - O2 Sensor: Stabilized Zirconia Sensor - Display: LCD Touch Screen - Exterior Material: Painted Steel (rear cover not painted) - Interior Material: Stainless Steel Copper Enriched Alloy - Insulation Material: Extruded polystyrene - Heating Method: Direct Heat & Air Jacket System 			
40	LABORATORY STEAM STERILIZER Quantity: 3			
	Laboratory, vertical, compact, floor-standing: <ul style="list-style-type: none"> - Electrically heated, microprocessor-controlled, programmable, automatic - temperature range: Max: 105°C - 138°C - Capacity: 150 L - Chamber and lid built in AISI 316L stainless steel. - External structure in AISI 304 stainless steel. - Manual top lid, hermetically closed by knobs and silicon gasket. - The entire control is made through the PLC (Programmable Logic Controller). - 3.5" color touch screen providing an easy and intuitive interface 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> - The heating is performed by electrical heaters. - Drying is performed through a vacuum caused by thermal shock over the entire height of the chamber, thus achieving a quick and effective drying. - As an option, drying can be performed by a vacuum pump. - A thermostatic steam trap is used to achieve automatic elimination of the steam condensates inside the chamber 			
41	CENTRIFUGE WITH COOLING Quantity: 3			
	<ul style="list-style-type: none"> -Max. speed: 18,000 rpm -Max. RCF: 23,542xg -Max. volume: 4 x 200ml -Speed range: 200 – 18,000 rpm -Temperature range: -20°C - 40°C -Running time: 59min 50s / 10s increments 99h 59min / 1min increments -Dimensions 40cm x 36cm x 70cm (WxHxD), -Weight 60 kg, -Power input: 660 W -Angle rotor 30x15ml Glass or 20x15ml conical, Angle: 35°, -Angle rotor 12x15ml, Angle: 32°, Angle: 45°, -Swing out rotor 2x3 Microtitreplates -Angle: 0-90° 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
42	MEDIA DOSING SYSTEM Quantity: 1			
	-Dosing range per dish 1 – 99.9 ml -Standard filling rate (up to 15 ml) circa 900 dishes/hour -Maximal dosing rate 500 ml/min -Dosing reproducibility circa 1% (at 15 ml)			
43	FAST POC qPCR SYSTEM Quantity: 1			
	-On-site molecular diagnostics -Portable, light -Two heating blocks, rechargeable lithium battery -LCD display with touch screen -No DNA/RNA extraction, isothermal amplification with real-time detection -Two- minutes cooking prepares the sample, test system for diagnostics of EHEC, EAggEC, EIEC/Shigella from bacterial culture and selective media			
44	WATER DESTILATION UNIT Quantity: 2			
	-Conductivity stage 1 /25 °C 2.3 µS/cm -Production output 12 L/h -Tank volume 24 L -Heater power max. 9 kW			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Power consumption max. 9 kW -Cooling water flow rate 3.3 L/min -Maximal pressure cooling water 7 bar 			
45	LABORATORY REFRIGERATOR Quantity: 3			
	<ul style="list-style-type: none"> -Temperature set range: + 4 ° C to + 15 ° C -Alarm set range: + 1 ° C to + 20 ° C -Gross volume: 1540 L -Interior dimensions: 1544 x 1291 x 707 mm (H x W x D) -External dimensions: 2050 x 1441 x 910 mm (H x W x D) -Safety door lock with 2 keys -Digital temperature indicator (Display: 0.1 Digits) -Controlled fan cooling system for constant temperature and even temperature across the entire refrigeration chamber. -Safety thermostat avoids the dropping of the stored product's temperature below + 2 ° C -Self contained alarm system with integrated battery (12V - 7 Ah) takes over the alarm function and temperature value measurements in case of power failure for at least 48 hours -Automatic fan switch off when the front door opens -Material interior body: Stainless steel -Material outer casing: Stainless steel with glass door 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	-Glass door heating (anti-mist) -12 Wire Shelf and 6 ST-Drawer -Noise level: 50 dB (A) -Energy consumption: 1.90 kWh / 24h -Heat ejection: 48 kcal / h -Hold Over time (from + 5 ° C to + 10 ° C): 1.42 h -Power: 280 W			
46	AUTOMATED IMMUNOASSAY SYSTEM Quantity: 2			
	-Compact, multiparametric automated immunoanalyzer -Technology ELFA (Enzyme Linked Fluorescent Assay) -Enables 10 different analytes to be used simultaneously and up to 30 results can be produced per hour -Used with assay kits covering: <ul style="list-style-type: none"> – hepatitis, – AIDS, – serology testing, – bacterial and viral antigen detection, – fertility/pregnancy monitoring, – thyroid hormones, – tumor markers, – endocrinology, – anemia, – sepsis 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> – immuno-hemostasis, – industrial microbiology -Mandatory tests in the range of clinical parameters: HIV DUO Ultra, Ultra HBs Ag, anti-HCV, Lyme IgM, IgG Lyme, EBV VCAIgM, EBV VCA / EA IgG EBV EBNA IgG, C. difficile Toxin A and B, C. difficile GDH -Mandatory tests for control of food pathogen detection: Salmonella, Listeria spp, Listeria monocytogenes, E.coli O157:H7, Campylobacter spp, Staphylococcal enterotoxins -Complete analysis on the reagents for a single test -Always ready for run, without startup and shutdown process -Ability to work individual analysis and series -Ability to work a number of different analyzes at the same time -Complete control of sample and reagent bar code -The possibility of networking 			
47	BACTERIAL COUNTING SYSTEM Quantity: 1			
	<ul style="list-style-type: none"> -Detects down to one organism per 100 mL -No dilutions (for counts up to 200 with Quanti-Tray; 2,419 with Quanti-Tray/2000; and up to 2,272 with Quanti-Tray/Legiolert). 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	-51-Well Trays -97-Well Trays			
48	DILUTOR Quantity: 1			
	-Fast and accurate dispensing: 12 sec. for 225ml / (9 sec in turbo mode) -Dilution accuracy: >99% -Turbo mode -Integrated bubble level -Motorized dispensing nozzle -Editable programs: 20 -Weighing range : 0.1 to 4000g -Stainless steel bag holder - extended lifetime -Watertight base - easy to clean -2 USB ports - 1 for upgrades / 1 for eTRACE connections -Plug and play removable pumps: 4 -Interactive display (color codes) -Dimensions (W x H x D): 275 x 250 x 400 mm -Weight: 8 kg -Electricity Supply: 100-240 V / 50-60 Hz / 120W			
49	LABORATORY TECHNICAL BALANCE Quantity: 2			
	- Range: 0-3000 g - Readability: 0.01 g			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> - Repeatability: 0.02 g - Linearity \pm 0.05 g - Cup diameter 120 mm 			
50	DRY STERILIZER Quantity: 2			
	<ul style="list-style-type: none"> - Minimal volume 120 L - Temperature range: +10 °C above ambient temperature to +300 °C -Up to 30% lower energy consumption compared to conventional units on the market -High temperature accuracy, APT line technology -Adjustable fan speed -Controller with LCD display and enhanced time functions -Electromechanical control of the exhaust air flap -2 chrome-plated racks -Class 2 integrated independent adjustable temperature safety device (DIN 12880 or equivalent) with visual alarm -Ergonomic handle design -Ethernet interface -USB port for recording data 			
51	LABORATORY PH METER Quantity: 3			
	<ul style="list-style-type: none"> -Parameter pH/mV/Temp -Measuring range pH -2.000 ... 19.999 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	-Measuring range Temperature -5.0 ... 105.0 °C -Measuring range mV -1200.0 ...+1200 / - 2000...+2000mV -Accuracy pH +/- 0.005/ +/-0.01/0.1 -Accuracy mV +/- 0.3 mV/ +/- 1 mV -Accuracy Temperature +/-0.1 K -Supplied with pH electrode			
52	THERMOSTAT DRY BLOCK SYSTEM WITH HEATING AND COOLING Quantity: 1			
	-Maximum load [kg] 0.3 -Permissible ON time [%] 100 -Timer yes -Timer display TFT -Time setting min. [s] 1 -Time setting max. [min] 6000 -Operating mode timer, continuous and program operation -Working with microtiter plates yes -Working temperature min. [°C] room temp. - 30 ±2 -Temperature setting range min. [°C] -10 -Heating temperature max. [°C] 110 -Set temperature resolution [°C] ±1 -Temperature control accuracy [°C] ±0.5 -Temperature stability [°C] 0.5 -Temperature homogeneity [°C] ±0.5 -Heating rate [°C/min] 5.5			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Temperature display yes -Cooling Rate [°C/min] 2 -Cooling Power [W] 72 embedded cooling unit yes -Number of exchangeable attachments 10 -Fixed safety circuit [°C] 145 -Permissible ambient temperature [°C] 5 - 40 -Permissible relative humidity [%] 80 -Protection class according to DIN EN 60529 IP 21 -For controlled dry heating of test tubes and microtubes in interchangeable aluminium blocks. -2 interchangeable blocks and 1 digital high-temperature system. Temperature control range to +120°C. -Fast heat-up time. -Digital time setting in 1 min increments. - Temperature stability and uniformity is maintained via sensitive, microprocessor control. -Built-in timer. -Included: <ul style="list-style-type: none"> - block removal tool - aluminium thermoblock for 2 ml tubes - aluminium thermoblock for microtubes 1.5 ml - Power cable set, 220/240 V. 			
53	THERMOSTAT DRY BLOCK SYSTEM WITH HEATING OPTION			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	Quantity: 1			
	<ul style="list-style-type: none"> -Number of blocks: 4 -Heat output [W] 412 -Heating temperature range [°C] room temp. +5° - 120 -Temperature display -Adjustment and display resolution [K] 1 -Connection for external temperature sensor CT (DIN12878) PT 1000 variation; DIN EN 60751 Kl. A [K] $\leq \pm (0.15 + 0.002 \times ITI)$ -Temperature stability within the blocks at 37°C [°C] ± 0.2 -Temperature stability within the blocks at 60°C [°C] ± 0.4 -Temperature Homogeneity @ 37°C [K] 0.2 -Temperature Homogeneity @ 60°C [K] 0.4 Heatingrate / Heat up time with external sensor [K/min] 5 -Set-up plate material Aluminium alloy -Set-up plate dimensions [mm] 96 x 304 -Fixed safety circuit [°C] 150 -Timer -Time setting range [min] 1 – 5999 -4 interchangeable blocks -Included: <ul style="list-style-type: none"> - block removal tool - 2 aluminium thermoblock for 2 ml tubes - 2 aluminium thermoblock for microtubes 1.5 ml 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
54	AUTOMATED SYSTEM FOR STAINING OF MICROORGANISMS Quantity: 1			
	<ul style="list-style-type: none"> -Standardization – slides are all stained the same way -Up to 30 slides ready in 5 minutes -Eco-friendly system -Staining adjustable to lab practices/habits -Standardized staining – innovative spray nozzles always dispense the same reagent volume -No cross contamination – each slide separated & fresh staining reagent used each time -Improved microorganism differentiation compared to manual and bath staining results -Full traceability of reagents, users and maintenance – troubleshoot problems & ensure compliance -User management to ensure standardization of the process -Clean and safe staining -Simply Load & Go (just 2 handling steps!) -Easy intuitive touch screen -Customize & program protocols -Rapid results (3-5 minutes) -Fully automated process from fixation to slide drying 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Improved lab workflow with single or batch slides possible -Can be integrated with Full Microbiology -- Lab Automation -Optional cytocentrifuge function 			
55	MICROBIAL DETECTION SYSTEM			
	<p>Quantity: 1</p> <ul style="list-style-type: none"> -Built-in control module and incubation chamber with optical detection system, where each position has its own optical detection system -Capacity of 120 bottles, with the possibility of subsequent upgrade capacity up to 840 (Modular system to increase easily the analytical capacity) -Colorimetric method of CO2 detection due to growth of microorganisms -Continuous monitoring of changes every 10 minutes, gives immediate notification of results with instant visual and audible alerts, including a remote alarm capability. -Possibility of adapting the capacity of the instrument for simultaneous detection of mycobacteria in one compartment of the incubator and other microorganisms in the other compartment (independent drawers for blood and SBFs samples and mycobacteria) 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<p>other than blood samples in the same instrument)</p> <ul style="list-style-type: none"> -Availability 24/7 -Software support for device handling, with the possibility of connecting to a LIS -Algorithms adapted to the detection of microorganisms from blood and other primarily sterile body fluids, sputum - validated by the manufacturer and approved by the FDA (CE-IVD marked and FDA clearance for blood and SBFs) -Algorithms adapted to the detection of microorganisms from bottles that stood at room temperature for 24 hours before investment without the risk of false negative results (Patented alghoritms for delayed bottles) -Customizable protocol incubation and protocol period by bottle -From the same bottle used, in addition to the detection of bacteria, the detection of fungi is also possible -Full solution of plastic bottles (unbreakble) to ensure maximum of safety - no glass bottles. -Storage up to 2000 results and growth curves -The possibility of archiving data to a USB drive -The ability to display and print different data and growth curves 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Bar code reader - built-in and external reader -Printer -UPS -Keyboard -Embedded touch scree 			
56	DENSITOMETER Quantity: 10			
	<ul style="list-style-type: none"> -Measurement range 0.00–15.00 McF -Display resolution 0.01 McF -Light source LED -Measurement wavelength (λ) $\lambda = 565 \pm 15$ nm -Accuracy (0.3–5.0 McF) $\pm 3\%$ -Measurement time: 1 s -Sample volume: not less than 2 ml -Tube external diameter: 12 mm, 16 mm (using A-12, A-16 adapter) or 18 mm (without adapter) -Possibility to restore factory calibration settings -Display: LCD -Dimensions (W×D×H): 165×115×75 mm -Weight: up to 0.7 kg -Input current/power consumption: 12 V, 7 mA / 0.1 W -External power supply: Input AC 100–240 V, 50/60 Hz; Output DC 12 V -Standard set: External power supply, A-16 			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
57	VORTEX MIXER Quantity: 12			
	-Type of movement orbital -Shaking stroke [mm] 4 -Permissible shaking weight (incl. attachment) [kg] 0.4 -Motor rating input [W] 39 -Motor rating output [W] 9 -Speed min (adjustable) [rpm] 500 -Speed max. [rpm] 2500 -Speed display scale -Speed control scale 0 - 6 -Operating mode continuous operation -Touch function yes -Dimensions (W x H x D) [mm] 120 x 140 x 138 -Weight [kg] 3.9 -Permissible ambient temperature [°C] 5 - 40 -Permissible relative humidity [%] 80 -Protection class according to DIN EN 60529 IP 21			
58	FLAMELESS STRILIZER Quantity: 6			
	-No combustion on the housing (Cool-Touch) -Sterilization temperature 750 ° C - 1000 ° C -Comfortable operation on touch or foot pedal (stainless steel foot pedal, cable length 3.0 m, included)			

1. Item Number	2. Specifications Required	3. Specifications Offered	4. Notes, remarks, ref to documentation	5. Evaluation Committee's notes
	<ul style="list-style-type: none"> -Energy efficient use through Thermocontol technology -Optical display of sterilization flow -IC sterilization source -Hallogen lamp -Sterilization of annealing tube before use -Quartz sterilization tube -Special quartz glass annealing tube -Very easy cleaning of devices and annealing tubes -Tube size external - minimum Ø 19 mm, internal - minimum Ø 16 mm, length maximum 112 mm -Fixed timer (10 seconds) automatically controls the heating of your instruments -Tested and certified for photobiological safety of lamps and lamp systems according to EN 62471 by BG ETEM - No risk to skin and eyes -Dimensions of the device maximum WxHxD 11 x 17 x 18 cm 			

Installation and after sales services for Lot No1

Specifications Required		Specifications Offered	Notes, remarks, ref to documentation	Evaluation Committee's notes YES/NO
Installation	Installation performed by contractor or authorised service provider; All the equipment must include all necessary parts and standards for its installation			
Testing	Testing of all basic function on a set of producer's standard samples commonly used for the corresponding instrument.			
Education	Theoretical education about basic functions of instrument, software and maintenance in Serbian or interpretation should be provided for 2 doctors and 3 technicians during 5 days performed by a manufacturer's licensed instructor no sooner than 1 month before installation.			
Start-up Training	Practical start-up training for 5 end users after installation and testing in all basic functions of the instrument on set of standard samples, commonly used for the corresponding instrument . Duration of training 1 day.			
Manuals	Instruction manual in English and Serbian			
Certificates and documentation	CE mark; Certificates conform to standards as specified in EU Directive 2002/98/EC; GMP (Good Manufacturing Practice)documentation; all in English language.			
Warranty	One year after provisional acceptance in accordance with the conditions laid down in Article 32 of the General Conditions.			

Annex II + III: Technical Specifications + Technical Offer - part II – Place of delivery/Acceptance

Item	ARTICLE	Name of person responsible for provisional and final acceptance	Place of acceptance
Lot 1 - REFERENCE MICROBIOLOGY LABORATORY			
Item 1-58	All items	Dr Verica Jovanović	Institute of Public Health of Serbia “Dr Milan Jovanovic Batut” Ministry of Health, Republic of Serbia Dr Subotica 5 11000 Belgrade

Authorised Contact Person	Delivery address
Name: Dr Verica Jovanović Phone: +381 11 2442741 Fax: N/A E-mail: verica_jovanovic@batut.org.rs Working hours: 7:30-15:30 (Monday to Friday)	Lot 1 Institute of Public Health of Serbia “Dr Milan Jovanovic Batut” Ministry of Health, Republic of Serbia Dr Subotica 5 11000 Belgrade